


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>UKAS CALIBRATION</p> <p>4068</p> <p>Accredited to ISO/IEC 17025:2005</p>	<h3>Q-Lab Corporation</h3> <p>Issue No: 002 Issue date: 13 December 2007</p>	
	<p>Stone Hill Road Farnworth Bolton BL4 9TP</p>	<p>Contact: Mr T Hindley Tel: +44 (0)1204 861616 Fax: +44 (0)1204 861617 E-Mail: info.eu@q.lab.com Website: www.q-lab.com</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks
RADIOMETRY Spectral responsivity	250 to 800 nm	5.8%	For calibration of Q-Lab type CR10 and CR20 radiometers only
END			